



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Organization of:

Aero Lab Solutions, LLC DBA Aero-Lab .
8361 NW 64th Street, Miami, FL 33166

*and hereby declares that the Organization is accredited in accordance with
the recognized International Standard:*

ISO/IEC 17025:2017

Whereby, technical competence has been confirmed for the associated scope supplement, in the fields of:

Chemical, Dimensional, Electrical and Mechanical Testing
(As detailed in the supplement)

Accreditation claims for conformity assessment activities shall only be made from the addresses referenced within this certificate and shall apply solely to those activities identified in the related scope. This Accreditation is granted subject to the Accreditation Body rules governing the Accreditation referred to above, and the Organization hereby commits to observing and complying with those rules in their entirety.

For PJLA:

Tracy Szerszen
President

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084

Initial Accreditation Date:

May 15, 2026

Issue Date:

May 15, 2026

Expiration Date:

August 31, 2028

Accreditation No.:

133894

Certificate No.:

L26-413

*The validity of this certificate is maintained through ongoing assessments based
on a continuous accreditation cycle. The validity of this certificate should be
confirmed through the PJLA website: www.pjilabs.com*



Certificate of Accreditation: Supplement

Aero Lab Solutions, LLC DBA Aero-Lab .

8361 NW 64th Street, Miami, FL 33166

Contact Name: Kristy Pantelakis Phone: 786-202-7580

Accreditation is granted to the facility to perform the following conformity assessment activities:

FIELD OF TEST	ITEMS, MATERIALS, OR PRODUCTS TESTED	COMPONENT, CHARACTERISTIC, PARAMETER TESTED	SPECIFICATION OR STANDARD METHOD	TECHNOLOGY OR TECHNIQUE USED	FLEX CODE	LOCATION OF ACTIVITY
Chemical	Penetrant	Water content percent in penetrant	ALQP10.1.3	Karl-Fischer Titrator	F1, F4	F
Chemical	Penetrant	Brightness in penetrant	ALQP10.1.4	S291 Photofluorometer	F1, F4	F
Chemical	Carrier II	Viscosity of Carrier II	ALQP10.1.8	Viscometer	F1, F4	F
Chemical	Emulsifier	Emulsifier concentration in percent	ALQP10.1.7, ALQP10.1.5	Karl-Fischer Titrator, Refractometer	F1, F4	F
Dimensional	TAM Panels	Dimensional measurement	ALQP10.1.1	B & W Vista Eye	F1	F
Electrical	Quick Break Tester	Quick Break Functionality	ALQP10.1.16	Magnetic Particle Bench	F1, F4	F
Mechanical	Metallic Materials	Adhesion	ALQP10.1.9	Visual Evaluation	F1	F
Mechanical	Metallic Materials	Salt Spray (Fog) Test	ALQP10.1.10	Environmental chamber unit	F1	F

1. Location of activity:

Location

F

Location

Conformity assessment activity is performed at the CABs fixed facility

2. Flex Code:

F0- Fixed scope item. No deviations allowed to the line item as identified, except for updating to the most recent version of an accredited standard method after verification.

F1- Laboratory has the capability to test a new item, material, matrix, or product similar in composition to item, material, matrix, or product identified on the scope

F2- Laboratory has the capability to introduce the newest revision of an accredited authoritative standard method (with no modifications) identified on the scope

F3- Laboratory has the capability to introduce a parameter/component/analyte to an accredited test method identified on the scope

F4- Laboratory has the capability to introduce a new revision of an accredited non-standard method using the same technology or technique identified on the scope

F5- Laboratory has the capability to introduce a validated method that is equivalent to an accredited method (using same technology or technique) identified on the scope

S